



P.O. Box 1089 Coldspring Tx 77331  
 Website: eastexlabs.com  
 Email: eastexlab@eastex.net  
 Tel: 936 653 3249



Texas A and M AgriLife  
 Texas A & M AgriLife  
 1335 Regents Park Dr. Ste. 260  
 Houston, TX 77058

**SM 2540 D - Quality Control**  
**Eastex Environmental Laboratory - Coldspring**

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
<b>Batch B0A4122 - No Prep</b>										
<b>Blank (B0A4122-BLK1)</b>				Prepared & Analyzed: 01/31/20						
TSS	ND	1.0	mg/L							
<b>Duplicate (B0A4122-DUP1)</b>				Source: C0A5857-01 Prepared & Analyzed: 01/31/20						
TSS	1350	1.0	mg/L		1380			2.20	10	
<b>Batch B0A4139 - No Prep Micro</b>										
<b>Blank (B0A4139-BLK1)</b>				Prepared & Analyzed: 01/30/20						
E coli IDEXX	ND	1	mpn/100ml							
<b>Duplicate (B0A4139-DUP1)</b>				Source: C0A5937-01 Prepared & Analyzed: 01/30/20						
E coli IDEXX	ND	2	mpn/100ml		ND				200	
<b>Batch B0B0016 - EPA 200.7</b>										
<b>Blank (B0B0016-BLK1)</b>				Prepared & Analyzed: 02/03/20						
Total Phosphorus	ND	0.0600	mg/L							
<b>Blank (B0B0016-BLK2)</b>				Prepared & Analyzed: 02/03/20						
Total Phosphorus	ND	0.0600	mg/L							
<b>Blank (B0B0016-BLK3)</b>				Prepared & Analyzed: 02/03/20						
Total Phosphorus	ND	0.0600	mg/L							
<b>Blank (B0B0016-BLK4)</b>				Prepared & Analyzed: 02/03/20						
Total Phosphorus	ND	0.0600	mg/L							
<b>LCS (B0B0016-BS1)</b>				Prepared & Analyzed: 02/03/20						
Total Phosphorus	2.62	0.0600	mg/L	2.52		104	85-115			
<b>LCS (B0B0016-BS2)</b>				Prepared & Analyzed: 02/03/20						
Total Phosphorus	2.59	0.0600	mg/L	2.52		103	85-115			
<b>LCS (B0B0016-BS3)</b>				Prepared & Analyzed: 02/03/20						
Total Phosphorus	2.68	0.0600	mg/L	2.52		106	85-115			
<b>LCS (B0B0016-BS4)</b>				Prepared & Analyzed: 02/03/20						
Total Phosphorus	2.71	0.0600	mg/L	2.52		107	85-115			
<b>LCS Dup (B0B0016-BSD1)</b>				Prepared & Analyzed: 02/03/20						
Total Phosphorus	2.58	0.0600	mg/L	2.52		102	85-115	1.78	20	
<b>LCS Dup (B0B0016-BSD2)</b>				Prepared & Analyzed: 02/03/20						
Total Phosphorus	2.58	0.0600	mg/L	2.52		102	85-115	0.351	20	
<b>LCS Dup (B0B0016-BSD3)</b>				Prepared & Analyzed: 02/03/20						
Total Phosphorus	2.72	0.0600	mg/L	2.52		108	85-115	1.62	20	

Eastex Environmental Laboratory - Coldspring

The results in this report apply to the samples analyzed in accordance with the chain of custody document.  
 This analytical report must be reproduced in its entirety. Alkalinity titrated to pH 4.5 endpoint.



P.O. Box 1089 Coldspring Tx 77331  
 Website: eastexlabs.com  
 Email: eastexlab@eastex.net  
 Tel: 936 653 3249



Texas A and M AgriLife  
 Texas A & M AgriLife  
 1335 Regents Park Dr. Ste. 260  
 Houston, TX 77058

**EPA 200.7 - Quality Control**  
**Eastex Environmental Laboratory - Coldspring**

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
<b>Batch B0B0016 - EPA 200.7</b>										
<b>LCS Dup (B0B0016-BSD4)</b>				Prepared & Analyzed: 02/03/20						
Total Phosphorus	2.73	0.0600	mg/L	2.52		108	85-115	0.821	20	
<b>Matrix Spike (B0B0016-MS1)</b>				Source: C0A0581-01 Prepared & Analyzed: 02/03/20						
Total Phosphorus	6.03	0.0600	mg/L	2.52	3.01	120	70-130			
<b>Matrix Spike (B0B0016-MS2)</b>				Source: C0A0581-09 Prepared & Analyzed: 02/03/20						
Total Phosphorus	3.59	0.0600	mg/L	2.52	0.552	120	70-130			
<b>Matrix Spike (B0B0016-MS3)</b>				Source: C0A4454-01 Prepared & Analyzed: 02/03/20						
Total Phosphorus	2.88	0.0600	mg/L	2.52	0.0433	113	70-130			
<b>Matrix Spike (B0B0016-MS4)</b>				Source: C0A6050-01 Prepared & Analyzed: 02/03/20						
Total Phosphorus	3.02	0.0600	mg/L	2.52	0.126	115	70-130			
<b>Batch B0B0163 - No Prep</b>										
<b>Blank (B0B0163-BLK1)</b>				Prepared & Analyzed: 02/04/20						
Nitrate+Nitrite as N	ND	0.02	mg/L							
<b>LCS (B0B0163-BS1)</b>				Prepared & Analyzed: 02/04/20						
Nitrate+Nitrite as N	0.980		mg/L	1.00		98.0	90-110			
<b>LCS Dup (B0B0163-BSD1)</b>				Prepared & Analyzed: 02/04/20						
Nitrate+Nitrite as N	1.01		mg/L	1.00		101	90-110	3.02	20	
<b>Matrix Spike (B0B0163-MS1)</b>				Source: C0A6047-01 Prepared & Analyzed: 02/04/20						
Nitrate+Nitrite as N	1.09	0.02	mg/L	1.00	0.0900	100	80-120			



P.O. Box 1089 Coldspring Tx 77331  
 Website: eastexlabs.com  
 Email: eastexlab@eastex.net  
 Tel: 936 653 3249



Texas A and M AgriLife  
 Texas A & M AgriLife  
 1335 Regents Park Dr. Ste. 260  
 Houston, TX 77058

**SM 4500 NH3 G - Quality Control**  
**Eastex Environmental Laboratory - Coldspring**

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
<b>Batch B0B1098 - No Prep</b>										
<b>Blank (B0B1098-BLK1)</b>				Prepared & Analyzed: 02/11/20						
Ammonia as N	ND	0.1	mg/L							
<b>LCS (B0B1098-BS1)</b>				Prepared & Analyzed: 02/11/20						
Ammonia as N	1.90		mg/L	2.00		95.0	90-110			
<b>LCS Dup (B0B1098-BSD1)</b>				Prepared & Analyzed: 02/11/20						
Ammonia as N	1.90		mg/L	2.00		95.0	90-110	0.00	20	
<b>Matrix Spike (B0B1098-MS1)</b>				Source: C0A6047-01 Prepared & Analyzed: 02/11/20						
Ammonia as N	2.8	0.1	mg/L	2.50	0.05	108	80-120			

*Mark Bourgeois*

Mark Bourgeois, Special Projects Manager

**Qualifiers**



**EASTEX ENVIRONMENTAL LABORATORY, INC.**

P.O. Box 1089 \* Colldspring, TX 77331 P.O. Box 631375 \* Nacogdoches, TX 75963-1375  
(936) 653-3249 \* (800) 525-0508 (936) 569-8879 \* FAX (936) 569-8951

White Copy-Follows Samples  
Yellow Copy-Laboratory  
Pink Copy-Client Copy

INVOICE TO: Christie Taylor www.eastexlabs.com

Company: TCWP Agrilik Extension  
Address: 1335 Regents Park 2nd  
Houston TX 77058  
Attn: Christie Taylor  
Phone#: 979-399-4009

REPORT TO: Christie Taylor  
Company: TCWP Agrilik Extension  
Address: 1335 Regents Park 2nd  
Houston TX 77058  
Attn: Christie Taylor  
Phone#: 979-399-4009  
Email: ctaylor@tamuedu  
P.O. #:

INSTRUCTIONS:  
C= Composite G= Grab  
Matrix: DW=Drinking Water WW=Wastewater SO=Soil/Sludge OT= Other  
Container Size: 1=Gallon 2=1/2 Gallon 3=Quart/Liter 4=500mL 5=250mL  
6=125mL (4oz) 7=60mL (2 oz) 8= 40mL Vial 9=Other

Sampler's Name (print):  
Christie Taylor  
Sampler's Signature:  
Christie Taylor  
Project Name:  
TCWP stormwater sampling

Type: P= Plastic G= Glass T= Teflon S= Sterile  
Preservatives: C=Chilled S=Sulfuric Acid N=Nitric Acid B=Base/Caustic Z= Zn Acetate  
ST=Sodium Thiosulfate H=HCL O= Other

Work Order ID  
A16155 EG 202-10  
Date  
1/30/2012

Time	Matrix	C or G	DO	pH	Cl2	Flow	Temp	Containers		
								#	Size	Type
							144	1	3 P	C
								1	5 P	N
								1	6 P/S	ST
								1	4 P	S/S
								1	4 P	S/S

Work Order ID	Sample ID	Date	Time	Matrix	C or G	DO	pH	Cl2	Flow	Temp	#	Size	Type	ANALYSIS REQUESTED			
														TSS	Total Phosphate	E. coli	
A16155	EG 202-10	1/30/2012	9:12	OT	G	7.8	7.43	-	-	144	1	3 P	C	X			
														X			
														X			
														X			
														X			

Relinquished By: Christie Taylor Date: 1/30/2012 Time: 10:20 Received Iced: YES / NO

Relinquished By: [Signature] Date: 1/30/2012 Time: 10:20 Received Iced: YES / NO

Relinquished By: [Signature] Date: 1/30/2012 Time: 10:20 Received Iced: YES / NO

LAP USE ONLY: Sample Condition Acceptable: YES / NO Temp: 2.8 Therm ID: 15 Logged In By: [Signature] Date: 1/30/2012 Time: 10:20 Received Iced: YES / NO