



P.O. Box 1089 Coldspring Tx 77331
 Website: eastexlabs.com
 Email: eastexlab@eastex.net
 Tel: 936 653 3249



Texas A and M AgriLife
 1335 Regents Park Dr. Ste. 260
 Houston TX, 77058

LABORATORY ANALYTICAL REPORT

Project: TX A and M Community Stormwater
 Sample Matrix: Water
 Client Matrix: Water

Sample Date & Time: 12/10/2019 12:49
 Collector: CT
 Sample Type: Grab
 Print Date: 12/23/2019

UTRP 101-3 C9L2887-01 (Water)

Analyte	Result	Reporting Limit	Units	Nelac Status	Batch	Analyzed Date & Time	Method	Notes
<u>Metals</u>								
Arsenic, Total	0.00151	0.000500	mg/L	A	B9L1864	12/17/2019 10:51	EPA 200.8	
Barium, Total	0.0237	0.00300	mg/L	A	B9L1864	12/17/2019 10:51	EPA 200.8	
Cadmium, Total	<0.00100	0.00100	mg/L	A	B9L1864	12/17/2019 10:51	EPA 200.8	
Chromium, Total	<0.00300	0.00300	mg/L	A	B9L1864	12/17/2019 10:51	EPA 200.8	
Lead, Total	0.00131	0.000500	mg/L	A	B9L1864	12/17/2019 10:51	EPA 200.8	
Mercury, Total	<0.000200	0.000200	mg/L	A	B9L1456	12/12/2019 12:38	EPA 245.1	
Selenium, Total	<0.00500	0.00500	mg/L	A	B9L1864	12/17/2019 10:51	EPA 200.8	
Silver, Total	<0.000500	0.000500	mg/L	A	B9L1864	12/17/2019 10:51	EPA 200.8	
Total Phosphorus	<0.0600	0.0600	mg/L	A	B9L2776	12/23/2019 15:04	EPA 200.7	

Microbiological Lab

E coli IDEXX	10	10	mpn/100ml	A	B9L1498	12/11/2019 16:40	Colilert 18	
--------------	----	----	-----------	---	---------	------------------	-------------	--

Organics

<u>TPH TX 1005</u>								
TPH (>C6 - C12)	<2.00	2.00	mg/L	A	B9L1602	12/13/2019 12:05	TX 1005	23
TPH (>C12 - C28)	<2.00	2.00	mg/L	A	B9L1602	12/13/2019 12:05	TX 1005	12, 53
TPH (>C28 - C35)	<5.00	5.00	mg/L	A	B9L1602	12/13/2019 12:05	TX 1005	12, 23, 53
Total TPH (>C6-C35)	<5.00	5.00	mg/L		B9L1602	12/13/2019 12:05	TX 1005	
Surrogate: 1-Chlorooctane			50.1 %	70-130	B9L1602	12/13/2019 12:05	TX 1005	
Surrogate: o-Terphenyl			35.4 %	70-130	B9L1602	12/13/2019 12:05	TX 1005	

Wet Lab

Ammonia as N	0.1	0.1	mg/L	A	B9L1899	12/16/2019 15:39	SM 4500 NH3 G	
--------------	-----	-----	------	---	---------	------------------	---------------	--

Eastex Environmental Laboratory - Coldspring

The results in this report apply to the samples analyzed in accordance with the chain of custody document.
 This analytical report must be reproduced in its entirety.

*NELAC Status: A=Accredited, N=Accreditation not offered, O=Not Accredited, P=Approved



P.O. Box 1089 Coldspring Tx 77331
 Website: eastexlabs.com
 Email: eastexlab@eastex.net
 Tel: 936 653 3249



Texas A and M AgriLife
 1335 Regents Park Dr. Ste. 260
 Houston TX, 77058

Project: TX A and M Community Stormwater
 Sample Matrix: Water
 Client Matrix: Water

Sample Date & Time: 12/10/2019 12:49
 Collector: CT
 Sample Type: Grab
 Print Date: 12/23/2019

UTRP 101-3
C9L2887-01 (Water)

Analyte	Result	Reporting Limit	Units	Nelac Status	Batch	Analyzed Date & Time	Method	Notes
<u>Wet Lab</u>								
Nitrate+Nitrite as N	0.77	0.02	mg/L	A	B9L1966	12/16/2019 08:30	SM 4500 NO3 F	
TSS	12.1	1.0	mg/L	A	B9L1571	12/12/2019 15:37	SM 2540 D	



P.O. Box 1089 Coldspring Tx 77331
 Website: eastexlabs.com
 Email: eastexlab@eastex.net
 Tel: 936 653 3249



Texas A and M AgriLife
 1335 Regents Park Dr. Ste. 260
 Houston TX, 77058

EPA 245.1 - Quality Control
Eastex Environmental Laboratory - Coldspring

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
Batch B9L1456 - EPA 245.1										
Blank (B9L1456-BLK1)				Prepared & Analyzed: 12/12/2019 11:44:17AM						
Mercury, Total	ND	0.000200	mg/L							
Blank (B9L1456-BLK2)				Prepared & Analyzed: 12/12/2019 12:16:35PM						
Mercury, Total	ND	0.000200	mg/L							
LCS (B9L1456-BS1)				Prepared & Analyzed: 12/12/2019 11:48:19AM						
Mercury, Total	0.00244	0.000200	mg/L	0.00250		97.6	80-120			
LCS (B9L1456-BS2)				Prepared & Analyzed: 12/12/2019 12:18:36PM						
Mercury, Total	0.00284	0.000200	mg/L	0.00250		114	80-120			
Leach Fluid Blank (B9L1456-LBK1)				Prepared & Analyzed: 12/12/2019 1:12:03PM						
Mercury, Total	ND	0.000200	mg/L							
Matrix Spike (B9L1456-MS1)		Source: C9L2121-01		Prepared & Analyzed: 12/12/2019 11:54:23AM						
Mercury, Total	0.00298	0.000200	mg/L	0.00250	0.000107	115	80-120			
Matrix Spike (B9L1456-MS2)		Source: C9L2479-01		Prepared & Analyzed: 12/12/2019 12:24:40PM						
Mercury, Total	0.00270	0.000200	mg/L	0.00250	0.000107	104	80-120			
Matrix Spike Dup (B9L1456-MSD1)		Source: C9L2121-01		Prepared & Analyzed: 12/12/2019 11:56:24AM						
Mercury, Total	0.00280	0.000200	mg/L	0.00250	0.000107	108	80-120	6.22	20	
Matrix Spike Dup (B9L1456-MSD2)		Source: C9L2479-01		Prepared & Analyzed: 12/12/2019 12:26:42PM						
Mercury, Total	0.00254	0.000200	mg/L	0.00250	0.000107	97.4	80-120	6.10	20	
Batch B9L1498 - No Prep Micro										
Blank (B9L1498-BLK1)				Prepared & Analyzed: 12/11/2019 4:40:00PM						
E coli IDEXX	ND		1 mpn/100ml							
Duplicate (B9L1498-DUP1)		Source: C9L2375-01		Prepared & Analyzed: 12/11/2019 4:40:00PM						
E coli IDEXX	ND		2 mpn/100ml		ND				200	
Batch B9L1571 - No Prep										
Blank (B9L1571-BLK1)				Prepared & Analyzed: 12/12/2019 3:37:00PM						
TSS	ND		1.0 mg/L							
Duplicate (B9L1571-DUP1)		Source: C9L2364-01		Prepared & Analyzed: 12/12/2019 3:37:00PM						
TSS	154	1.0	mg/L		152			1.31	10	

Eastex Environmental Laboratory - Coldspring

*The results in this report apply to the samples analyzed in accordance with the chain of custody document.
 This analytical report must be reproduced in its entirety.*

*NELAC Status: A=Accredited, N=Accreditation not offered, O=Not Accredited, P=Approved



P.O. Box 1089 Coldspring Tx 77331
 Website: eastexlabs.com
 Email: eastexlab@eastex.net
 Tel: 936 653 3249



Texas A and M AgriLife
 1335 Regents Park Dr. Ste. 260
 Houston TX, 77058

SM 2540 D - Quality Control
Eastex Environmental Laboratory - Coldspring

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC %REC	Limits	RPD	RPD Limit	Notes
---------	--------	-----------------	-------	-------------	---------------	-----------	--------	-----	-----------	-------

Batch B9L1571 - No Prep

Batch B9L1602 - No Prep- Organic

Blank (B9L1602-BLK1)

Prepared & Analyzed: 12/12/2019 3:40:00PM

TPH (>C6 - C12)	ND	2.00	mg/L							
TPH (>C12 - C28)	ND	2.00	mg/L							
TPH (>C28 - C35)	ND	5.00	mg/L							
Total TPH (>C6-C35)	ND	5.00	mg/L							
Surrogate: 1-Chlorooctane	27.0		mg/L	33.3		80.9	70-130			
Surrogate: o-Terphenyl	23.1		mg/L	33.3		69.4	70-130			52

LCS (B9L1602-BS1)

Prepared & Analyzed: 12/12/2019 4:25:00PM

TPH (>C6 - C12)	28.9	2.00	mg/L	33.3		86.7	70-130			
TPH (>C12 - C28)	30.4	2.00	mg/L	33.3		91.4	70-130			
TPH (>C28 - C35)	53.8	5.00	mg/L	66.7		80.7	70-130			
Surrogate: 1-Chlorooctane	29.8		mg/L	33.3		89.5	70-130			
Surrogate: o-Terphenyl	26.0		mg/L	33.3		78.1	70-130			

Matrix Spike (B9L1602-MS1)

Source: C9L2902-01

Prepared & Analyzed: 12/12/2019 5:56:00PM

TPH (>C6 - C12)	29.1	2.00	mg/L	33.3	ND	87.2	70-130			
TPH (>C12 - C28)	32.3	2.00	mg/L	33.3	ND	96.9	70-130			12
TPH (>C28 - C35)	43.5	5.00	mg/L	66.7	ND	65.2	70-130			12, 23
Surrogate: 1-Chlorooctane	29.6		mg/L	33.3		88.7	70-130			
Surrogate: o-Terphenyl	26.1		mg/L	33.3		78.3	70-130			

Matrix Spike Dup (B9L1602-MSD1)

Source: C9L2902-01

Prepared & Analyzed: 12/12/2019 6:41:00PM

TPH (>C6 - C12)	22.8	2.00	mg/L	33.3	ND	68.3	70-130	24.3	25	23
TPH (>C12 - C28)	24.1	2.00	mg/L	33.3	ND	72.2	70-130	29.2	25	12, 53
TPH (>C28 - C35)	26.0	5.00	mg/L	66.7	ND	39.1	70-130	50.1	25	12, 23, 53
Surrogate: 1-Chlorooctane	20.6		mg/L	33.3		62.0	70-130			52
Surrogate: o-Terphenyl	17.4		mg/L	33.3		52.2	70-130			52

Batch B9L1864 - EPA 200.8

Blank (B9L1864-BLK1)

Prepared & Analyzed: 12/16/2019 5:39:03PM

Arsenic, Total	ND	0.000500	mg/L							
Barium, Total	ND	0.00300	mg/L							
Cadmium, Total	ND	0.00100	mg/L							
Chromium, Total	ND	0.00300	mg/L							

Eastex Environmental Laboratory - Coldspring

*The results in this report apply to the samples analyzed in accordance with the chain of custody document.
 This analytical report must be reproduced in its entirety.*

*NELAC Status: A=Accredited, N=Accreditation not offered, O=Not Accredited, P=Approved



P.O. Box 1089 Coldspring Tx 77331
 Website: eastexlabs.com
 Email: eastexlab@eastex.net
 Tel: 936 653 3249



Texas A and M AgriLife
 1335 Regents Park Dr. Ste. 260
 Houston TX, 77058

EPA 200.8 - Quality Control
Eastex Environmental Laboratory - Coldspring

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch B9L1864 - EPA 200.8

Blank (B9L1864-BLK1)

Prepared & Analyzed: 12/16/2019 5:39:03PM

Lead, Total	ND	0.000500	mg/L							
Selenium, Total	ND	0.00500	mg/L							
Silver, Total	ND	0.000500	mg/L							

Blank (B9L1864-BLK2)

Prepared & Analyzed: 12/16/2019 7:10:14PM

Lead, Total	ND	0.000500	mg/L							
-------------	----	----------	------	--	--	--	--	--	--	--

LCS (B9L1864-BS1)

Prepared & Analyzed: 12/16/2019 5:48:16PM

Arsenic, Total	0.101	0.000500	mg/L	0.100		101	85-115			
Barium, Total	0.102	0.00300	mg/L	0.100		102	85-115			
Cadmium, Total	0.100	0.00100	mg/L	0.100		100	85-115			
Chromium, Total	0.0971	0.00300	mg/L	0.100		97.1	85-115			
Lead, Total	0.102	0.000500	mg/L	0.100		102	85-115			
Selenium, Total	0.0994	0.00500	mg/L	0.100		99.4	85-115			
Silver, Total	0.102	0.000500	mg/L	0.100		102	85-115			

LCS (B9L1864-BS2)

Prepared & Analyzed: 12/16/2019 7:14:50PM

Lead, Total	0.0954	0.000500	mg/L	0.100		95.4	85-115			
-------------	--------	----------	------	-------	--	------	--------	--	--	--

Matrix Spike (B9L1864-MS1)

Source: C9L2243-01

Prepared & Analyzed: 12/16/2019 6:01:51PM

Arsenic, Total	0.0794	0.000500	mg/L	0.100	0.00152	77.8	70-130			
Barium, Total	0.123	0.00300	mg/L	0.100	0.0487	74.5	70-130			
Cadmium, Total	0.0767	0.00100	mg/L	0.100	ND	76.7	70-130			
Chromium, Total	0.0748	0.00300	mg/L	0.100	0.000231	74.5	70-130			
Lead, Total	0.0784	0.000500	mg/L	0.100	0.000127	78.3	70-130			
Selenium, Total	0.0785	0.00500	mg/L	0.100	ND	78.5	70-130			
Silver, Total	0.0779	0.000500	mg/L	0.100	ND	77.9	70-130			

Matrix Spike (B9L1864-MS2)

Source: C9L2351-01

Prepared & Analyzed: 12/16/2019 7:28:27PM

Lead, Total	0.0839	0.000500	mg/L	0.100	0.000360	83.6	70-130			
-------------	--------	----------	------	-------	----------	------	--------	--	--	--

Matrix Spike Dup (B9L1864-MSD1)

Source: C9L2243-01

Prepared & Analyzed: 12/16/2019 6:06:16PM

Arsenic, Total	0.0908	0.000500	mg/L	0.100	0.00152	89.2	70-130	13.4	20	
Barium, Total	0.140	0.00300	mg/L	0.100	0.0487	91.4	70-130	12.9	20	
Cadmium, Total	0.0878	0.00100	mg/L	0.100	ND	87.8	70-130	13.4	20	
Chromium, Total	0.0867	0.00300	mg/L	0.100	0.000231	86.5	70-130	14.8	20	
Lead, Total	0.0902	0.000500	mg/L	0.100	0.000127	90.1	70-130	14.0	20	
Selenium, Total	0.0876	0.00500	mg/L	0.100	ND	87.6	70-130	10.9	20	

Eastex Environmental Laboratory - Coldspring

*The results in this report apply to the samples analyzed in accordance with the chain of custody document.
 This analytical report must be reproduced in its entirety.*

*NELAC Status: A=Accredited, N=Accreditation not offered, O=Not Accredited, P=Approved



P.O. Box 1089 Coldspring Tx 77331
 Website: eastexlabs.com
 Email: eastexlab@eastex.net
 Tel: 936 653 3249



Texas A and M AgriLife
 1335 Regents Park Dr. Ste. 260
 Houston TX, 77058

EPA 200.8 - Quality Control
Eastex Environmental Laboratory - Coldspring

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
Batch B9L1864 - EPA 200.8										
Matrix Spike Dup (B9L1864-MSD1)		Source: C9L2243-01			Prepared & Analyzed: 12/16/2019 6:06:16PM					
Silver, Total	0.0901	0.000500	mg/L	0.100	ND	90.1	70-130	14.6	20	
Matrix Spike Dup (B9L1864-MSD2)		Source: C9L2351-01			Prepared & Analyzed: 12/16/2019 7:32:52PM					
Lead, Total	0.0833	0.000500	mg/L	0.100	0.000360	82.9	70-130	0.746	20	
Batch B9L1899 - No Prep										
Blank (B9L1899-BLK1)					Prepared & Analyzed: 12/16/2019 3:32:00PM					
Ammonia as N	ND	0.1	mg/L							
LCS (B9L1899-BS1)					Prepared & Analyzed: 12/16/2019 3:28:00PM					
Ammonia as N	2.19		mg/L	2.00		110	90-110			
LCS Dup (B9L1899-BSD1)					Prepared & Analyzed: 12/16/2019 3:26:00PM					
Ammonia as N	2.18		mg/L	2.00		109	90-110	0.548	20	
MRL Check (B9L1899-MRL1)					Prepared & Analyzed: 12/17/2019 6:29:00AM					
Ammonia as N	0.121		mg/L	0.100		121	0-200			
Matrix Spike (B9L1899-MS1)		Source: C9L2885-01			Prepared & Analyzed: 12/16/2019 3:34:00PM					
Ammonia as N	2.7	0.1	mg/L	2.50	0.1	102	80-120			
Batch B9L1966 - No Prep										
Blank (B9L1966-BLK1)					Prepared & Analyzed: 12/16/2019 8:30:00AM					
Nitrate+Nitrite as N	ND	0.02	mg/L							
LCS (B9L1966-BS1)					Prepared & Analyzed: 12/16/2019 8:30:00AM					
Nitrate+Nitrite as N	1.04		mg/L	1.00		104	90-110			
LCS Dup (B9L1966-BSD1)					Prepared & Analyzed: 12/16/2019 8:30:00AM					
Nitrate+Nitrite as N	1.04		mg/L	1.00		104	90-110	0.00	20	
MRL Check (B9L1966-MRL1)					Prepared & Analyzed: 12/16/2019 8:30:00AM					
Nitrate+Nitrite as N	0.0190		mg/L	0.0200		95.0	0-200			
Matrix Spike (B9L1966-MS1)		Source: C9L1828-02			Prepared & Analyzed: 12/16/2019 8:30:00AM					
Nitrate+Nitrite as N	1.12	0.02	mg/L	1.00	0.0200	110	80-120			
Batch B9L2776 - EPA 200.7										

Eastex Environmental Laboratory - Coldspring

*The results in this report apply to the samples analyzed in accordance with the chain of custody document.
 This analytical report must be reproduced in its entirety.*

*NELAC Status: A=Accredited, N=Accreditation not offered, O=Not Accredited, P=Approved



P.O. Box 1089 Coldspring Tx 77331
 Website: eastexlabs.com
 Email: eastexlab@eastex.net
 Tel: 936 653 3249



Texas A and M AgriLife
 1335 Regents Park Dr. Ste. 260
 Houston TX, 77058

EPA 200.7 - Quality Control
Eastex Environmental Laboratory - Coldspring

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
Batch B9L2776 - EPA 200.7										
Blank (B9L2776-BLK1)					Prepared & Analyzed: 12/23/2019 1:55:04PM					
Total Phosphorus	ND	0.0600	mg/L							
Blank (B9L2776-BLK2)					Prepared & Analyzed: 12/23/2019 2:54:08PM					
Total Phosphorus	ND	0.0600	mg/L							
LCS (B9L2776-BS1)					Prepared & Analyzed: 12/23/2019 3:36:19PM					
Total Phosphorus	2.41	0.0600	mg/L	2.52		95.8	85-115			
LCS (B9L2776-BS2)					Prepared & Analyzed: 12/23/2019 3:46:43PM					
Total Phosphorus	2.40	0.0600	mg/L	2.52		95.2	85-115			
LCS Dup (B9L2776-BSD1)					Prepared & Analyzed: 12/23/2019 3:39:46PM					
Total Phosphorus	2.41	0.0600	mg/L	2.52		95.8	85-115	0.0282	20	
LCS Dup (B9L2776-BSD2)					Prepared & Analyzed: 12/23/2019 3:50:11PM					
Total Phosphorus	2.54	0.0600	mg/L	2.52		101	85-115	5.86	20	
MRL Check (B9L2776-MRL1)					Prepared & Analyzed: 12/23/2019 3:32:49PM					
Total Phosphorus	0.0651		mg/L	0.0600		109	0-200			
Matrix Spike (B9L2776-MS1)					Source: C9L1828-01 Prepared & Analyzed: 12/23/2019 2:12:27PM					
Total Phosphorus	2.32	0.0600	mg/L	2.52	0.0777	88.8	70-130			
Matrix Spike (B9L2776-MS2)					Source: C9L2887-01 Prepared & Analyzed: 12/23/2019 3:11:30PM					
Total Phosphorus	2.42	0.0600	mg/L	2.52	0.0385	94.5	70-130			



P.O. Box 1089 Coldspring Tx 77331
Website: eastexlabs.com
Email: eastexlab@eastex.net
Tel: 936 653 3249



Texas A and M AgriLife
1335 Regents Park Dr. Ste. 260
Houston TX, 77058

Notes and Definitions

- 53 RPD Recovery outside acceptance limits due to matrix interference.
- 52 Surrogate recovery outside of acceptance limits due to matrix interference.
- 23 Spike recovery outside of acceptance limits due to matrix interference.
- 12 CCV associated with sample batch did not meet acceptance criteria.
- DET Analyte DETECTED
- ND Analyte NOT DETECTED at or above the reporting limit
- NR Not Reported
- dry Sample results reported on a dry weight basis
- RPD Relative Percent Difference



EASTEX ENVIRONMENTAL LABORATORY, INC.

P.O. Box 1089 * Coldspring, TX 77331 P.O. Box 631375 * Nacogdoches, TX 75963-1375
 (936) 653-3249 * (800) 525-0508 (936) 569-8879 * FAX (936) 569-8951

White Copy-Follows Samples
 Yellow Copy-Laboratory
 Pink Copy-Client Copy

REPORT TO: Christie Taylor
Company: AgriLife Ext. Srv.
Address: 1335 Regents Park Dr
Houston TX 77058
Attn: Christie Taylor
Phone#: 979-399-4009
Email: ctaylor@tamu.edu
P.O. #: _____

INVOICE TO: Christie Taylor
Company: AgriLife Ext. Srv.
Address: 1335 Regents Park Dr
Houston TX 77058
Attn: Christie Taylor
Phone#: 979-399-4009
INSTRUCTIONS: _____
C or G: C= Composite G= Grab
Matrix: DW=Drinking Water WW=Wastewater SO=Soil/Sludge OT=Other
Container Size: 1=Gallon 2=1/2 Gallon 3=Quart/Liter 4=500ml 5=250ml
 6=125mL (4oz) 7=60mL (2 oz) 8= 40mL Vial 9=Other
Type: P= Plastic G= Glass T= Teflon S= Sterile
Preservatives: C=Chilled S=Sulfuric Acid N=Nitric Acid B=Base/Caustic Z= Zn Acetate
 ST=Sodium Thiosulfate H=HCL O= Other

Project Name: ICWP Stormwater Sampling
Sampler's Name (print): Christie Taylor
Sampler's Signature: Christie Taylor
Project Name: ICWP Stormwater Sampling

Work Order ID	Sample ID	Date	Field Data				Containers			ANALYSIS REQUESTED				
			Time	Matrix	C or G	DO	pH	CI2	Flow		Temp	#	Size	Type
2887	UTRP 101-3	12-10-19	12:49	OT	C	10.3	9.88	15°C	1	3	P	C	X	TSS
									1	5	P/S	C/N	X	Total Phosphate
									1	6	P	C/ST	X	E.coli
									1	4	P	C/SA	X	NH ₃ N
									1	4	P	C/SA	X	NO ₂ + NO ₃
									1	4	P	C/N	X	RCA metals
									3	8	P/S	C/A	X	TPH

Relinquished By: Christie Taylor **Date:** 12-11-19 **Time:** 1400
Received By: _____ **Date:** _____ **Time:** _____
Relinquished By: _____ **Date:** 12-11-19 **Time:** 1400
Received By: _____ **Date:** _____ **Time:** _____
LAB USE ONLY **Sample Condition Acceptable:** YES / NO **Temp C:** 2.0 **Logged In By:** _____
Alternate Check In: _____ **Date:** _____ **Time:** _____

*Thermometer has 0.0 factor and recorded temperature is actual temperature