



P.O. Box 1089 Coldspring Tx 77331
Website: eastexlabs.com
Email: eastexlab@eastex.net
Tel: 936 653 3249




August 10, 2020

Texas A&M Agrilife
Texas A and M AgriLife
1335 Regents Park Dr. Ste. 260
Houston, TX 77058

TX A and M Community Stormwater

Enclosed are the results of analyses for samples received by the laboratory on 07/22/20 14:50, with Lab ID Number C0G4866. If you have any questions concerning this report, please feel free to contact me.

Sincerely,


Mark Bourgeois
Special Projects Manager



P.O. Box 1089 Coldspring Tx 77331
 Website: eastexlabs.com
 Email: eastexlab@eastex.net
 Tel: 936 653 3249



Texas A and M AgriLife
 1335 Regents Park Dr. Ste. 260
 Houston TX, 77058

LABORATORY ANALYTICAL REPORT

Project: TX A and M Community Stormwater
 Sample Matrix: Water
 Client Matrix: Water

Sample Date & Time: 07/22/2020 10:39
 Collector: CT
 Sample Type: Grab
 Print Date: 8/10/2020

**PTWB 302-5
 C0G4866-01 (Water)**

Analyte	Result	Reporting Limit	Units	Nelac Status	Batch	Analyzed Date & Time	Method	Notes
<u>Metals</u>								
Arsenic, Total	0.00213	0.000500	mg/L	A	B0G3324	07/31/2020 03:29	EPA 200.8	
Barium, Total	0.0495	0.00300	mg/L	A	B0G3324	07/31/2020 03:29	EPA 200.8	
Cadmium, Total	<0.00100	0.00100	mg/L	A	B0G3324	07/31/2020 03:29	EPA 200.8	
Chromium, Total	0.00334	0.00300	mg/L	A	B0G3324	07/31/2020 03:29	EPA 200.8	
Lead, Total	<0.000500	0.000500	mg/L	A	B0G3324	07/31/2020 03:29	EPA 200.8	
Mercury, Total	<0.000200	0.000200	mg/L	A	B0G3166	07/23/2020 15:37	EPA 245.1	
Selenium, Total	<0.00500	0.00500	mg/L	A	B0G3324	07/31/2020 03:29	EPA 200.8	
Silver, Total	<0.000500	0.000500	mg/L	A	B0G3324	07/31/2020 03:29	EPA 200.8	
Total Phosphorus	<0.0600	0.0600	mg/L	A	B0H0141	08/07/2020 09:45	EPA 200.7	
<u>Microbiological Lab</u>								
E coli IDEXX	<10	10	mpn/100ml	A	B0G3196	07/22/2020 15:24	Colilert 18	

Organics

<u>TPH TX 1005</u>								
TPH (>C6 - C12)	<2.00	2.00	mg/L	A	B0G3492	08/06/2020 16:04	TX 1005	
TPH (>C12 - C28)	<2.00	2.00	mg/L	A	B0G3492	08/06/2020 16:04	TX 1005	
TPH (>C28 - C35)	<5.00	5.00	mg/L	A	B0G3492	08/06/2020 16:04	TX 1005	
Total TPH (>C6-C35)	<5.00	5.00	mg/L		B0G3492	08/06/2020 16:04	TX 1005	
Surrogate: 1-Chlorooctane			111 %	70-130	B0G3492	08/06/2020 16:04	TX 1005	
Surrogate: o-Terphenyl			114 %	70-130	B0G3492	08/06/2020 16:04	TX 1005	

Wet Lab

Ammonia as N	<0.1	0.1	mg/L	A	B0G3518	07/27/2020 14:34	SM 4500 NH3 G	
Nitrate+Nitrite as N	0.02	0.02	mg/L	A	B0H0045	08/03/2020 13:06	SM 4500 NO3 F	
TSS	3.9	1.0	mg/L	A	B0G3120	07/23/2020 13:27	SM 2540 D	25

Eastex Environmental Laboratory - Coldspring

The results in this report apply to the samples analyzed in accordance with the chain of custody document.
 This analytical report must be reproduced in its entirety.

*NELAC Status: A=Accredited, N=Accreditation not offered, O=Not Accredited, P=Approved



P.O. Box 1089 Coldspring Tx 77331
 Website: eastexlabs.com
 Email: eastexlab@eastex.net
 Tel: 936 653 3249



Texas A and M AgriLife
 1335 Regents Park Dr. Ste. 260
 Houston TX, 77058

SM 2540 D - Quality Control
Eastex Environmental Laboratory - Coldspring

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
Batch B0G3120 - No Prep										
Blank (B0G3120-BLK1)				Prepared & Analyzed: 7/23/2020 1:27:00PM						
TSS	ND	1.0	mg/L							
Duplicate (B0G3120-DUP1)				Source: C0G4445-01 Prepared & Analyzed: 7/23/2020 1:27:00PM						
TSS	254	1.0	mg/L		252			0.791	10	
Batch B0G3166 - EPA 245.1										
Blank (B0G3166-BLK1)				Prepared & Analyzed: 7/23/2020 3:19:26PM						
Mercury, Total	ND	0.000200	mg/L							
LCS (B0G3166-BS1)				Prepared & Analyzed: 7/23/2020 3:21:27PM						
Mercury, Total	0.00251	0.000200	mg/L	0.00250		100	75-115			
Matrix Spike (B0G3166-MS1)				Source: C0G4380-01 Prepared & Analyzed: 7/23/2020 3:29:32PM						
Mercury, Total	0.00305	0.000200	mg/L	0.00250	ND	122	70-130			
Matrix Spike Dup (B0G3166-MSD1)				Source: C0G4380-01 Prepared & Analyzed: 7/23/2020 3:31:33PM						
Mercury, Total	0.00309	0.000200	mg/L	0.00250	ND	124	70-130	1.30	20	
Batch B0G3196 - No Prep Micro										
Blank (B0G3196-BLK1)				Prepared & Analyzed: 7/22/2020 3:24:00PM						
E coli IDEXX	ND	1	mpn/100ml							
Duplicate (B0G3196-DUP1)				Source: C0G4454-01 Prepared & Analyzed: 7/22/2020 3:24:00PM						
E coli IDEXX	ND	2	mpn/100ml		ND				200	
Batch B0G3324 - EPA 200.8										
Blank (B0G3324-BLK1)				Prepared & Analyzed: 7/31/2020 12:24:09AM						
Arsenic, Total	ND	0.000500	mg/L							
Barium, Total	ND	0.00300	mg/L							
Cadmium, Total	ND	0.00100	mg/L							
Chromium, Total	ND	0.00300	mg/L							
Lead, Total	ND	0.000500	mg/L							
Selenium, Total	ND	0.00500	mg/L							
Silver, Total	ND	0.000500	mg/L							
LCS (B0G3324-BS1)				Prepared & Analyzed: 7/31/2020 12:29:22AM						
Arsenic, Total	0.101	0.000500	mg/L	0.100		101	85-115			

Eastex Environmental Laboratory - Coldspring

*The results in this report apply to the samples analyzed in accordance with the chain of custody document.
 This analytical report must be reproduced in its entirety.*

*NELAC Status: A=Accredited, N=Accreditation not offered, O=Not Accredited, P=Approved



P.O. Box 1089 Coldspring Tx 77331
 Website: eastexlabs.com
 Email: eastexlab@eastex.net
 Tel: 936 653 3249



Texas A and M AgriLife
 1335 Regents Park Dr. Ste. 260
 Houston TX, 77058

EPA 200.8 - Quality Control
Eastex Environmental Laboratory - Coldspring

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch B0G3324 - EPA 200.8

LCS (B0G3324-BS1)

Prepared & Analyzed: 7/31/2020 12:29:22AM

Barium, Total	0.109	0.00300	mg/L	0.100		109	85-115			
Cadmium, Total	0.0983	0.00100	mg/L	0.100		98.3	85-115			
Chromium, Total	0.0970	0.00300	mg/L	0.100		97.0	85-115			
Lead, Total	0.108	0.000500	mg/L	0.100		108	85-115			
Selenium, Total	0.0967	0.00500	mg/L	0.100		96.7	85-115			
Silver, Total	0.107	0.000500	mg/L	0.100		107	85-115			

Matrix Spike (B0G3324-MS1)

Source: C0G4520-01

Prepared & Analyzed: 7/31/2020 12:44:47AM

Arsenic, Total	0.0989	0.000500	mg/L	0.100	0.00166	97.2	70-130			
Barium, Total	0.109	0.00300	mg/L	0.100	0.00704	102	70-130			
Cadmium, Total	0.0950	0.00100	mg/L	0.100	0.000978	94.1	70-130			
Chromium, Total	0.0944	0.00300	mg/L	0.100	ND	94.4	70-130			
Lead, Total	0.104	0.000500	mg/L	0.100	ND	104	70-130			
Selenium, Total	0.0939	0.00500	mg/L	0.100	ND	93.9	70-130			
Silver, Total	0.101	0.000500	mg/L	0.100	ND	101	70-130			

Matrix Spike Dup (B0G3324-MSD1)

Source: C0G4520-01

Prepared & Analyzed: 7/31/2020 12:49:53AM

Arsenic, Total	0.0995	0.000500	mg/L	0.100	0.00166	97.9	70-130	0.703	20	
Barium, Total	0.105	0.00300	mg/L	0.100	0.00704	97.6	70-130	4.26	20	
Cadmium, Total	0.0957	0.00100	mg/L	0.100	0.000978	94.7	70-130	0.679	20	
Chromium, Total	0.0964	0.00300	mg/L	0.100	ND	96.4	70-130	2.04	20	
Lead, Total	0.0977	0.000500	mg/L	0.100	ND	97.7	70-130	5.86	20	
Selenium, Total	0.0958	0.00500	mg/L	0.100	ND	95.8	70-130	1.97	20	
Silver, Total	0.102	0.000500	mg/L	0.100	ND	102	70-130	1.20	20	

Batch B0G3492 - No Prep- Organic

Blank (B0G3492-BLK1)

Prepared & Analyzed: 8/6/2020 12:01:00PM

TPH (>C6 - C12)	ND	2.00	mg/L							
TPH (>C12 - C28)	ND	2.00	mg/L							
TPH (>C28 - C35)	ND	5.00	mg/L							
Total TPH (>C6-C35)	ND	5.00	mg/L							
Surrogate: 1-Chlorooctane	27.6		mg/L	33.3		82.8	70-130			
Surrogate: o-Terphenyl	28.4		mg/L	33.3		85.1	70-130			

LCS (B0G3492-BS1)

Prepared & Analyzed: 8/6/2020 2:33:00PM

TPH (>C6 - C12)	30.1	2.00	mg/L	33.3		90.4	70-130			
TPH (>C12 - C28)	30.6	2.00	mg/L	33.3		91.7	70-130			

Eastex Environmental Laboratory - Coldspring

*The results in this report apply to the samples analyzed in accordance with the chain of custody document.
 This analytical report must be reproduced in its entirety.*

*NELAC Status: A=Accredited, N=Accreditation not offered, O=Not Accredited, P=Approved



P.O. Box 1089 Coldspring Tx 77331
 Website: eastexlabs.com
 Email: eastexlab@eastex.net
 Tel: 936 653 3249



Texas A and M AgriLife
 1335 Regents Park Dr. Ste. 260
 Houston TX, 77058

TX 1005 - Quality Control
Eastex Environmental Laboratory - Coldspring

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Batch B0G3492 - No Prep- Organic

LCS (B0G3492-BS1)										
										Prepared & Analyzed: 8/6/2020 2:33:00PM
TPH (>C28 - C35)	51.4	5.00	mg/L	66.7		77.2	70-130			
Surrogate: 1-Chlorooctane	32.0		mg/L	33.3		95.9	70-130			
Surrogate: o-Terphenyl	32.2		mg/L	33.3		96.5	70-130			
Matrix Spike (B0G3492-MS1)										
										Source: C0G3573-01
										Prepared & Analyzed: 8/6/2020 6:20:00PM
TPH (>C6 - C12)	40.1	2.00	mg/L	33.3	5.53	104	70-130			
TPH (>C12 - C28)	37.7	2.00	mg/L	33.3	2.99	104	70-130			
TPH (>C28 - C35)	59.6	5.00	mg/L	66.7	ND	89.4	70-130			
Surrogate: 1-Chlorooctane	36.6		mg/L	33.3		110	70-130			
Surrogate: o-Terphenyl	37.3		mg/L	33.3		112	70-130			
Matrix Spike Dup (B0G3492-MSD1)										
										Source: C0G3573-01
										Prepared & Analyzed: 8/6/2020 7:06:00PM
TPH (>C6 - C12)	40.4	2.00	mg/L	33.3	5.53	105	70-130	0.676	25	
TPH (>C12 - C28)	39.2	2.00	mg/L	33.3	2.99	109	70-130	4.08	25	
TPH (>C28 - C35)	57.3	5.00	mg/L	66.7	ND	86.0	70-130	3.88	25	
Surrogate: 1-Chlorooctane	36.8		mg/L	33.3		110	70-130			
Surrogate: o-Terphenyl	37.5		mg/L	33.3		113	70-130			

Batch B0G3518 - No Prep

Blank (B0G3518-BLK1)										
										Prepared & Analyzed: 7/27/2020 2:21:00PM
Ammonia as N	ND	0.1	mg/L							
LCS (B0G3518-BS1)										
										Prepared & Analyzed: 7/27/2020 2:18:00PM
Ammonia as N	2.00		mg/L	2.00		99.8	90-110			
LCS Dup (B0G3518-BSD1)										
										Prepared & Analyzed: 7/27/2020 2:16:00PM
Ammonia as N	2.05		mg/L	2.00		103	90-110	2.81	20	
MRL Check (B0G3518-MRL1)										
										Prepared & Analyzed: 7/27/2020 2:19:00PM
Ammonia as N	0.121		mg/L	0.100		121	0-200			
Matrix Spike (B0G3518-MS1)										
										Source: C0G4538-04
										Prepared & Analyzed: 7/27/2020 2:23:00PM
Ammonia as N	2.7	0.1	mg/L	2.50	0.1	103	80-120			

Batch B0H0045 - No Prep

Blank (B0H0045-BLK1)										
										Prepared & Analyzed: 8/3/2020 12:38:00PM

Eastex Environmental Laboratory - Coldspring The results in this report apply to the samples analyzed in accordance with the chain of custody document.

This analytical report must be reproduced in its entirety.

*NELAC Status: A=Accredited, N=Accreditation not offered, O=Not Accredited, P=Approved



P.O. Box 1089 Coldspring Tx 77331
 Website: eastexlabs.com
 Email: eastexlab@eastex.net
 Tel: 936 653 3249



Texas A and M AgriLife
 1335 Regents Park Dr. Ste. 260
 Houston TX, 77058

SM 4500 NO3 F - Quality Control
Eastex Environmental Laboratory - Coldspring

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
Batch B0H0045 - No Prep										
Blank (B0H0045-BLK1)				Prepared & Analyzed: 8/3/2020 12:38:00PM						
Nitrate+Nitrite as N	ND	0.02	mg/L							
LCS (B0H0045-BS1)				Prepared & Analyzed: 8/3/2020 12:41:00PM						
Nitrate+Nitrite as N	1.05		mg/L	1.00		105	90-110			
Matrix Spike (B0H0045-MS1)				Source: C0G3279-01		Prepared & Analyzed: 8/3/2020 12:43:00PM				
Nitrate+Nitrite as N	1.51	0.02	mg/L	1.00	0.432	108	80-120			
Matrix Spike Dup (B0H0045-MSD1)				Source: C0G3279-01		Prepared & Analyzed: 8/3/2020 12:46:00PM				
Nitrate+Nitrite as N	1.53	0.02	mg/L	1.00	0.432	110	80-120	1.32	20	
Batch B0H0141 - EPA 200.7										
Blank (B0H0141-BLK1)				Prepared & Analyzed: 8/7/2020 9:00:09AM						
Total Phosphorus	ND	0.0600	mg/L							
Blank (B0H0141-BLK2)				Prepared & Analyzed: 8/7/2020 9:59:16AM						
Total Phosphorus	ND	0.0600	mg/L							
LCS (B0H0141-BS1)				Prepared & Analyzed: 8/7/2020 9:03:37AM						
Total Phosphorus	2.36	0.0600	mg/L	2.52		93.7	85-115			
LCS (B0H0141-BS2)				Prepared & Analyzed: 8/7/2020 10:02:45AM						
Total Phosphorus	2.45	0.0600	mg/L	2.52		97.2	85-115			
LCS Dup (B0H0141-BSD1)				Prepared & Analyzed: 8/7/2020 9:07:05AM						
Total Phosphorus	2.45	0.0600	mg/L	2.52		97.1	85-115	3.58	20	
LCS Dup (B0H0141-BSD2)				Prepared & Analyzed: 8/7/2020 10:06:15AM						
Total Phosphorus	2.53	0.0600	mg/L	2.52		100	85-115	3.06	20	
MRL Check (B0H0141-MRL1)				Prepared & Analyzed: 8/7/2020 1:01:19PM						
Total Phosphorus	0.0718		mg/L	0.0600		120	0-200			
Matrix Spike (B0H0141-MS1)				Source: C0G4538-01		Prepared & Analyzed: 8/7/2020 9:17:29AM				
Total Phosphorus	2.56	0.0600	mg/L	2.52	0.0517	99.4	70-130			
Matrix Spike (B0H0141-MS2)				Source: C0G5000-02		Prepared & Analyzed: 8/7/2020 10:16:41AM				
Total Phosphorus	3.02	0.0600	mg/L	2.52	0.127	115	70-130			

Eastex Environmental Laboratory - Coldspring

*The results in this report apply to the samples analyzed in accordance with the chain of custody document.
 This analytical report must be reproduced in its entirety.*

*NELAC Status: A=Accredited, N=Accreditation not offered, O=Not Accredited, P=Approved



P.O. Box 1089 Coldspring Tx 77331
Website: eastexlabs.com
Email: eastexlab@eastex.net
Tel: 936 653 3249



Texas A and M AgriLife
1335 Regents Park Dr. Ste. 260
Houston TX, 77058

Notes and Definitions

- 25 Residue not in the method range of 2.5-200 mg.
- DET Analyte DETECTED
- ND Analyte NOT DETECTED at or above the reporting limit
- NR Not Reported
- dry Sample results reported on a dry weight basis
- RPD Relative Percent Difference



EASTEX ENVIRONMENTAL LABORATORY, INC.

P.O. Box 1089 * Coldspring, TX 77331 (936) 653-3249 * (800) 525-0508
P.O. Box 631375 * Nacogdoches, TX 75963-1375 (936) 569-8879 * FAX (936) 569-8951

White Copy-Follows Samples
Yellow Copy-Laboratory
Pink Copy-Client Copy

INVOICE TO: Christie Taylor
www.eastexlabs.com

Company: AgriLife Extension
Address: 1335 Regents Park 260
Houston TX 77058
Attn:
Phone#: 979-399-4009

Remarks:
RCRA metals
NDA + NO3
NH3 W
Total Phosphate
E. coli
TPH

INSTRUCTIONS:
C or G: C= Composite G= Grab
Matrix: DW=Drinking Water WW=Wastewater SO=Soil/Sludge OT= Other
Container Size: 1=1/2 Gallon 3=Quart/Liter 4=500mL 5=250mL
6=125mL (4oz) 7=60mL (2 oz) 8=40mL Vial 9=Other
Type: P= Plastic G= Glass T= Teflon S= Sterile
Preservatives: C=Chilled S=Sulfuric Acid N=Nitric Acid B=Base/Caustic Z= Zn Acetate
ST=Sodium Thiosulfate H=HCL O= Other

ANALYSIS REQUESTED
TS

Field Data

Time	Matrix	C or G	DO	pH	CI2	Flow	Temp	#	Size	Type	Pres
10:31	W	G	8.4	8.67	-	-	29.6	1	3	P	C
								1	4	P	S
								2	5	P	S
								2	5	P	S
								1	5	P	C/N
								1	6	P/S	S/T
								3	8	P/S	C/A

Work Order ID	Sample ID	Date
U8604	PTWB 302-5	7/22

Relinquished By: Christie Taylor Date: 7/22/20 Time: 11:35 Received Iced: YES / NO

Relinquished By: W Smith Date: 7/22/20 Time: 14:50 Received Iced: YES / NO

Relinquished By: W Smith Date: 7/22/20 Time: 14:50 Received Iced: YES / NO

LAB USE ONLY
Sample Condition Acceptable: YES / NO
Alternate Check In: _____ Date: _____
Temp C: 14 Logged In By: W Smith Therm ID: 15
Date: 7/22/20 Time: 14:50

*Thermometer has 0.0 factor and recorded temperature is actual temperature