



P.O. Box 1089 Coldspring Tx 77331
 Website: eastexlabs.com
 Email: eastexlab@eastex.net
 Tel: 936 653 3249



Texas A and M AgriLife
 Texas A & M AgriLife
 1250 Bay Area Blvd
 Houston, TX 77058

LABORATORY ANALYTICAL REPORT

Project: Tx Coastal Exploration Green

Sample Site:	EG-P3-1	Sample Number:	C3D3479-01		Collector:					
Sample Type:	Composite				Sampled:	04/11/2023	17:25			
Sample Matrix:	Water				Received:	04/12/2023	16:40			
Client Matrix:	Water									
Analyte	Result	Reporting Limit	Units	Nelac Status	Batch	Analyzed	Analyst	Method	Notes	
Ammonia as N	<0.1	0.1	mg/L	A	B3D1852	04/17/2023 09:57	SAC	SM 4500 NH3 G		
Total Phosphorus	<0.0600	0.0600	mg/L	A	B3D2620	04/19/2023 15:54	KJH	EPA 200.7		
TSS	<4.0	4.0	mg/L	A	B3D1722	04/13/2023 14:07	WLS	SM 2540 D		

CORRECTED REPORT
ORIGINAL REPORT
DATE 4/20/23
project name



P.O. Box 1089 Coldspring Tx 77331
 Website: eastexlabs.com
 Email: eastexlab@eastex.net
 Tel: 936 653 3249



Texas A and M AgriLife
 Texas A & M AgriLife
 1250 Bay Area Blvd
 Houston, TX 77058

SM 2540 D - Quality Control
Eastex Environmental Laboratory - Coldspring

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
Batch B3D1722 - No Prep										
Blank (B3D1722-BLK1)				Prepared & Analyzed: 04/13/23						
TSS	ND	1.0	mg/L							
Duplicate (B3D1722-DUP1)				Source: C3D2746-01 Prepared & Analyzed: 04/13/23						
TSS	92.0	1.0	mg/L		90.0			2.20	10	
Batch B3D1852 - No Prep										
Blank (B3D1852-BLK1)				Prepared & Analyzed: 04/17/23						
Ammonia as N	ND	0.1	mg/L							
LCS (B3D1852-BS1)				Prepared & Analyzed: 04/17/23						
Ammonia as N	2.08		mg/L	2.00		104	90-110			
Matrix Spike (B3D1852-MS1)				Source: C3D2676-01 Prepared & Analyzed: 04/17/23						
Ammonia as N	8.8	0.1	mg/L	2.50	6.6	89.6	80-120			
Matrix Spike Dup (B3D1852-MSD1)				Source: C3D2676-01 Prepared & Analyzed: 04/17/23						
Ammonia as N	8.7	0.1	mg/L	2.50	6.6	85.0	80-120	1.31	20	
Batch B3D2620 - EPA 200.7										
Blank (B3D2620-BLK1)				Prepared & Analyzed: 04/19/23						
Total Phosphorus	ND	0.0600	mg/L							
LCS (B3D2620-BS1)				Prepared & Analyzed: 04/19/23						
Total Phosphorus	2.89	0.0600	mg/L	2.52		115	85-115			
Matrix Spike (B3D2620-MS1)				Source: C3D2611-01 Prepared & Analyzed: 04/19/23						
Total Phosphorus	7.38	0.0600	mg/L	2.52	4.67	107	70-130			
Matrix Spike Dup (B3D2620-MSD1)				Source: C3D2611-01 Prepared & Analyzed: 04/19/23						
Total Phosphorus	7.45	0.0600	mg/L	2.52	4.67	110	70-130	1.01	20	

Mark Bourgeois

Mark Bourgeois, Special Projects Manager

Qualifiers



EASTEX ENVIRONMENTAL LABORATORY, INC.
 P. O. Box 1089 • Colleyville, TX 77331 | P. O. Box 631375 • Nacogdoches, TX 75963-1375
 (846) 525-0508 • FAX 1936 653-3172 | (936) 564-5679 • FAX 1936) 569-8951
 www.eastexlab.com

REPORT TO:

INVOICE TO:

1 Company **AgriLife Extension**
 Address **1335 Regent Park Dr**
Houston, TX 77057
 Attn. **Christie Taylor**
 Phone # **979-599-4007** **832-794-5985**
 Fax #

2 Company
 Address
 Attn.
 Phone #
 Fax #
 P. O. #

3 Project Name **Water Quality Baseline**
 4 Sample ID **EG-Water**
 5 Date **4/12/23**
 6 Time **14:33**
 7 Operator **RE**

8 Analytes
AMMONIA
TOTAL PHOSPHATE

C3D3479

Order ID	Sample ID	D		T		C		O		P		M		S		Z		P		C
		1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	
EG-P3-1	1																			

Received by: **Christie Taylor** Date: **4/12/23** Time: **14:33**
 Received by: **Christie Taylor** Date: **4/12/23** Time: **16:40**
 Chain of Custody: **Paulson Z 032417**